



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

1 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/1000	1.5 %
2	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/3000	1.2 %
3	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/500	1.57 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

2 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
4	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1500 (Part 2):2021, ISO 6506-2:2017 and ASTM E10:2023	HBW 2.5/187.5	1.5 %
5	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 2.5/62.5	1.5 %
6	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 5/250	1.59 %
7	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017, ASTM E10 : 2023	HBW 5/750	1.53 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

3 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
8	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Reference Test Block	Using Standardizing Hardness Testing Machine as per IS 1500 (Part 3) : 2019, ISO 6506-3 : 20214 and ASTM E10 : 2023	HBW 10/3000	1.1 %
9	MECHANICAL-HARDNESS TESTING MACHINES	Depth Measuring System of Rockwell Hardness Tester	as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	Upto 250 µm	3 µm
10	MECHANICAL-HARDNESS TESTING MACHINES	Diagonal Measuring System of Vickers and Knoop Hardness Tester	Using Glass Scale as per IS 6885 (Part 2 & 3) : 2020, ISO 4545-2 & 3 : 2020 ASTM E92 : 2023, ASTM E384 : 2022, IS 1501 (Part 2 & 3) : 2020 and ISO 6507-2 & 3 : 2023	Upto 1.2 mm	0.86 %
11	MECHANICAL-HARDNESS TESTING MACHINES	Indentation Diameter Measuring System of Brinell Hardness Tester	Using Glass Scale as per IS 1500 (Part 2 & 3) : 2021, ISO 6506-2 & 3 : 2017 and ASTM E10 : 2023	Upto 5 mm	0.31 %
12	MECHANICAL-HARDNESS TESTING MACHINES	Indirect verification of Rockwell ball indenter	Using Standardising Hardness Testing Machine as per IS 1586 (Part 2):2018, ISO 6508-2:2023 and ASTM E18:2024	HRBW	0.5 HRBW





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

4 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
13	MECHANICAL-HARDNESS TESTING MACHINES	Indirect Verification of Rockwell Diamond Indenter (ACDN)	Using Standardising Hardness Testing Machine as per IS 1586 (Part 2):2018. ISO 6508-2:2023 and ASTM E18:2024	HRA, HRC, HRD, HRN	0.5 HR
14	MECHANICAL-HARDNESS TESTING MACHINES	Knoop Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 6885 (Part 2):2020, ISO 4545-2:2017, ASTM E92:2023 and ASTM E384:2022	HK 0.2	3.7 %
15	MECHANICAL-HARDNESS TESTING MACHINES	Knoop Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 6885 (Part 2):2020, ISO 4545-2:20217, ASTM E92:2023 and ASTM E384:2022	HK 0.5	3.5 %
16	MECHANICAL-HARDNESS TESTING MACHINES	Knoop Reference Test Block	Using Standardising hardness testing machine as per IS 6885 (Part 3):2020, ISO 4545-3:2017, ASTM E92:2023 and ASTM E384:2022	HK 0.2	5.0 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

5 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
17	MECHANICAL- HARDNESS TESTING MACHINES	Knoop Reference Test Block	Using Standardising hardness testing machine as per IS 6885 (Part 3):2020, ISO 4545-3:20217, ASTM E92:2023 and ASTM E384:2022	HK 0.5	4.0 %
18	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.05	6.5 %
19	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:20218 ASTM E92:2023 and ASTM E384:2022	HV 0.1	5.5 %
20	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.2	5.6 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

6 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
21	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.3	5.0 %
22	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.5	3.5 %
23	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 1	2.5 %
24	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3):2020. ISO 6507-3:2018, ASTM E92:2023 and ASTM 384:2022	HV 1	2.3 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

7 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018, ASTM E92 : 2023 and ASTM 384 : 2022	HV0.05	6.05 %
26	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3):2020. ISO 6507-3:2018, ASTM E92:2023 and ASTM 384:2022	HV0.1	5.5 %
27	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018, ASTM E92 : 2023 and ASTM 384 : 2022	HV0.2	5.64 %
28	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3):2020, ISO 6507-3 : 2018, ASTM E92 : 2023 and ASTM 384 : 2022	HV0.3	4.47 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

8 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
29	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018, ASTM E92 : 2023 and ASTM 384 : 2022	HV0.5	3.21 %
30	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR15N	0.6 HR15N
31	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR15TW	0.8 HR15TW





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

9 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
32	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR30N	0.6 HR30N
33	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR30TW	0.8 HR30TW
34	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR45N	0.6 HR45N



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

10 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
35	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR45TW	0.8 HR45TW
36	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRA	0.6 HRA
37	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRBW	0.94 HRBW



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

11 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
38	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRC	0.52 HRC
39	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3):2018. ISO 6508-3:2023 and ASTM E18:2024	HR15N	0.4 HR15N
40	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3) : 2018, ISO 6508-3 : 2023 and ASTM E18 : 2024	HR15TW	0.93 HR15TW
41	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3) : 2018, ISO 6508-3 : 2023 and ASTM E18 : 2024	HR30N	0.52 HR30N





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

12 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3) : 2018, ISO 6508-3 : 2023 and ASTM E18 : 2024	HR30TW	0.94 HR30TW
43	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3) : 2018, ISO 6508-3 : 2023 and ASTM E18 : 2024	HR45N	0.6 HR45N
44	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3):2018. ISO 6508-3:2023 and ASTM E18:2024	HR45TW	0.9 HR45TW
45	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3):2018. ISO 6508-3:2023 and ASTM E18:2024	HRA	0.4 HRA



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

13 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
46	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3):2018. ISO 6508-3:2023 and ASTM E18:2024	HRBW	0.58 HRBW
47	MECHANICAL- HARDNESS TESTING MACHINES	Rockwell Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1586 (Part 3):2018. ISO 6508-3:2023 and ASTM E18:2024	HRC	0.4 HRC
48	MECHANICAL- HARDNESS TESTING MACHINES	Test Force of Brinell Hardness Tester	Using Load Cells as per IS 1500 (Part 2 and Part 3) : 2019, ISO 6506-2 & 3 : 2017 and ASTM E10 : 2023	612.9 N to 29420 N	0.21 %
49	MECHANICAL- HARDNESS TESTING MACHINES	Test Force of Rockwell Hardness Tester	Using Load Cells as per IS 1586 (Part 2 and Part 3) : 2018, ISO 6508-2 & 3 : 2023 and ASTM E18 : 2024	29.42 N to 1471 N	0.13 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

14 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
50	MECHANICAL-HARDNESS TESTING MACHINES	Test Force of Vickers and Knoop Hardness Tester	Using Load Cells as per IS 6885 (Part 2 & 3) : 2020, ISO 4545-2 & 3 : 2017 ASTM E92 : 2023 ASTM E384 : 2022, IS 1501 (Part 2 & 3):2020 and ISO 6507- 2 & 3 : 2018	0.4903 N to 490.3 N	0.25 %
51	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 10	1.3 %
52	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 20	1.3 %
53	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 30	1.3 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

15 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
54	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 5	1.6 %
55	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 50	1.3 %
56	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018 and ASTM E92 : 2023	HV 10	1.02 %
57	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3):2020. ISO 6507-3:2018 and ASTM E92:2023	HV 20	1.1 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

16 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
58	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018 and ASTM E92 : 2023	HV 30	1.08 %
59	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018 and ASTM E92 : 2023	HV 5	1.56 %
60	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Reference Test Block	Using Standardising Hardness Testing Machine as per IS 1501 (Part 3) : 2020, ISO 6507-3 : 2018 and ASTM E92 : 2023	HV 50	1.07 %
61	MECHANICAL-IMPACT TESTING MACHINE	Charpy Impact Testing Machine	Using Inclinator, Load Cell, Gauges by Direct Method and Indirect Method (for verification only) as per ISO 148-2 : 2016, IS 1757 (Part 2) : 2020 and ASTM E23 : 2024	Up to 500 J	1.03 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

17 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
62	MECHANICAL- IMPACT TESTING MACHINE	Izod Impact Testing Machine	Using Inclinator, Load Cell, Gauges by Direct Method as per BS 131 (Part 4) : 1972, IS 1598 : 2024 and ASTM E23 : 2024	Up to 170 J	0.8 %
63	MECHANICAL- UTM, TENSION CREEP AND TORSION TESTING MACHINE	Extensometer (L. C. : 0.1 µm or coarser)	Using Extensometer Calibration Fixture, Electronic Probe and DRO as per IS 12872:2021, ISO 9513:2012 and ASTM E83 : 2023	Up to 10 mm	0.005 mm





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

18 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/1000	1.5 %
2	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/3000	1.2 %
3	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 10/500	1.57 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

19 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
4	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1500 (Part 2):2021, ISO 6506-2:2017 and ASTM E10:2023	HBW 2.5/187.5	1.5 %
5	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 2.5/62.5	1.5 %
6	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017 and ASTM E10 : 2023	HBW 5/250	1.59 %
7	MECHANICAL-HARDNESS TESTING MACHINES	Brinell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1500 (Part 2) : 2021, ISO 6506-2 : 2017, ASTM E10 : 2023	HBW 5/750	1.53 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

20 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
8	MECHANICAL-HARDNESS TESTING MACHINES	Depth Measuring System of Rockwell Hardness Tester	as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	Upto 250 $\mu$ m	3 $\mu$ m
9	MECHANICAL-HARDNESS TESTING MACHINES	Diagonal Measuring System of Vickers and Knoop Hardness Tester	Using Glass Scale as per IS 6885 (Part 2 & 3) : 2020, ISO 4545-2 & 3 : 2020 ASTM E92 : 2023, ASTM E384 : 2022, IS 1501 (Part 2 & 3) : 2020 and ISO 6507-2 & 3 : 2023	Upto 1.2 mm	0.86 %
10	MECHANICAL-HARDNESS TESTING MACHINES	Indentation Diameter Measuring System of Brinell Hardness Tester	Using Glass Scale as per IS 1500 (Part 2 & 3) : 2021, ISO 6506-2 & 3 : 2017 and ASTM E10 : 2023	Upto 5 mm	0.31 %
11	MECHANICAL-HARDNESS TESTING MACHINES	Knoop Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 6885 (Part 2):2020, ISO 4545-2:2017, ASTM E92:2023 and ASTM E384:2022	HK 0.2	3.7 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

21 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	MECHANICAL-HARDNESS TESTING MACHINES	Knoop Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 6885 (Part 2):2020, ISO 4545-2:20217, ASTM E92:2023 and ASTM E384:2022	HK 0.5	3.5 %
13	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.05	6.5 %
14	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:20218 ASTM E92:2023 and ASTM E384:2022	HV 0.1	5.5 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

22 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.2	5.6 %
16	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.3	5.0 %
17	MECHANICAL- HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 0.5	3.5 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

23 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
18	MECHANICAL-HARDNESS TESTING MACHINES	Micro-Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 ASTM E92:2023 and ASTM E384:2022	HV 1	2.5 %
19	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR15N	0.6 HR15N
20	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR15TW	0.8 HR15TW





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

24 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
21	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR30N	0.6 HR30N
22	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR30TW	0.8 HR30TW
23	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR45N	0.6 HR45N



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

25 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
24	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HR45TW	0.8 HR45TW
25	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRA	0.6 HRA
26	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRBW	0.94 HRBW



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

26 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
27	MECHANICAL-HARDNESS TESTING MACHINES	Rockwell Hardness Tester	Using Standard Hardness Test Blocks by Indirect Method as per IS 1586 (Part 2) : 2018, ISO 6508-2 : 2023 and ASTM E18 : 2024	HRC	0.52 HRC
28	MECHANICAL-HARDNESS TESTING MACHINES	Test Force of Brinell Hardness Tester	Using Load Cells as per IS 1500 (Part 2 and Part 3) : 2019, ISO 6506-2 & 3 : 2017 and ASTM E10 : 2023	612.9 N to 29420 N	0.21 %
29	MECHANICAL-HARDNESS TESTING MACHINES	Test Force of Rockwell Hardness Tester	Using Load Cells as per IS 1586 (Part 2 and Part 3) : 2018, ISO 6508-2 & 3 : 2023 and ASTM E18 : 2024	29.42 N to 1471 N	0.13 %
30	MECHANICAL-HARDNESS TESTING MACHINES	Test Force of Vickers and Knoop Hardness Tester	Using Load Cells as per IS 6885 (Part 2 & 3) : 2020, ISO 4545-2 & 3 : 2017 ASTM E92 : 2023 ASTM E384 : 2022, IS 1501 (Part 2 & 3):2020 and ISO 6507- 2 & 3 : 2018	0.4903 N to 490.3 N	0.25 %





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

27 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
31	MECHANICAL- HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 10	1.3 %
32	MECHANICAL- HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 20	1.3 %
33	MECHANICAL- HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 30	1.3 %
34	MECHANICAL- HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 5	1.6 %



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

28 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
35	MECHANICAL-HARDNESS TESTING MACHINES	Vickers Hardness Tester	Using Standard Hardness test blocks by indirect method as per IS 1501 (Part 2):2020, ISO 6507-2:2018 and ASTM E92:2023	HV 50	1.3 %
36	MECHANICAL-IMPACT TESTING MACHINE	Charpy Impact Testing Machine	Using Inclinator, Load Cell, Gauges by Direct Method and Indirect Method (for verification only) as per ISO 148-2 : 2016, IS 1757 (Part 2) : 2020 and ASTM E23 : 2024	Up to 500 J	1.03 %
37	MECHANICAL-IMPACT TESTING MACHINE	Izod Impact Testing Machine	Using Inclinator, Load Cell, Gauges by Direct Method as per BS 131 (Part 4) : 1972, IS 1598 : 2024 and ASTM E23 : 2024	Up to 170 J	0.8 %
38	MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE	Extensometer (L. C. : 0.1 µm or coarser)	Using Extensometer Calibration Fixture, Electronic Probe and DRO as per IS 12872:2021, ISO 9513:2012 and ASTM E83 : 2023	Up to 10 mm	0.005 mm



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

FINE (MIRASHI) CALIBRATION AND TESTING LABORATORIES LLP, FINE  
MANUFACTURING INDUSTRIES, B-7/12, MIDC AREA, MIRAJ, SANGLI,  
MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4429

**Page No**

29 of 29

**Validity**

03/06/2025 to 02/06/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
39	MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE	Force Calibration of Uni-axial Testing Machine (UTM, CTM) - Compression Mode	Using Force Proving Instruments as per IS 1828 (Part 1) : 2022, ISO 7500-1 : 2018 and ASTM E4 : 2024	100 N to 1000 kN	0.35 %
40	MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE	Uni-axial Testing Machine (UTM, TTM) - Tension mode	Using Force Proving Instruments as per IS 1828 (Part 1):2022, ISO 7500-1 : 2022 and ASTM E4 : 2024	20 N to 300 kN	0.35 %

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.